

DRAFT
AMENDMENT

TC 8
(CH)

SUBMITTED
FOR CIML
BALLOT

**Draft Amendment to R 138:
Vessels for commercial transactions**

*Draft submitted for CIML postal ballot on 2009.03.09
Voting closes on 2009.06.01. The final Draft will be
submitted for approval to the CIML at its 44th Meeting*

Draft Amendment to R 138



ORGANISATION INTERNATIONALE
DE MÉTROLOGIE LÉGALE

INTERNATIONAL ORGANIZATION
OF LEGAL METROLOGY

1 Introduction

OIML R 138:2007 was approved by the CIML during its 42nd Meeting held in Shanghai. In the meantime it was decided to amend OIML R 138 to take into consideration late comments.

2 Amendment (2009) to OIML R 138:2007 (E)

2.1 Section 2.6

2.6 Conventional true capacity

a) For measuring container bottles and casks: Contained volume of water at a reference temperature with an uncertainty appropriate for the given purpose.

b) For capacity serving measures: Delivered volume of water at a reference temperature with an uncertainty appropriate for the given purpose.

2.2 Section 5.2.3

5.2.3 Measuring container bottles

In addition to the requirements of 5.2.1, measuring container bottles shall be marked with one of the following marks, either on the lower rim or the bottom:

- when filled to a constant level: the distance (expressed in millimeters, followed by the symbol mm) from the brim to the filling level corresponding to the nominal capacity; or
- when filled to a constant ullage: the number (expressed in millilitres (ml) or centilitres (cl) not followed by the symbol ml/cl) that is equal to the brim capacity.

2.3 Annex A

Annex A is an informative Annex.